



## CERTIFICATE

No. U8 001888 0002 Rev. 01

Holder of Certificate:

Amico Clinical Solutions, Corp

55 East Wilmot St

Richmond Hill ON L4B 1A3

CANADA

Production Facility(ies): 095914

Certification Mark:



Product:

**General Medical Devices** 

Surgical Light

Model(s):

Roll stand, Permanent, Amico iCE30m 160

Parameters:

Roll stand, Permanent:

Rated voltage: Rated frequency: 100-240 Vac 50/60 Hz 4-2 A

Rated input current: Protection class:

Degree of protection:

No applied part.

Amico iCE30m 160:

Rated voltage:

24 Vdc

Rated input power

200 W

Degree of protection:

No applied part.

Tested

ANSI/AAMI ES60601-1:2005/A1:2012-08 excl Biocompatibility (Cl. 11.7), PEMS (CI. 14) and EMC (CI. 17)

according to:

CAN/CSA-C22.2 No. 60601-1:2014 excluding clauses 11.7, 14 and

CAN/CSA-C22.2 No. 60601-2-41:2011/A1:2015-11

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This productcertification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined inISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜDAmerica Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certificationbody.

Test report no.:

7169003892-100

Date,

2019-03-14

(Peter Keith)

